



NATIONAL
ENVIRONMENTAL
TESTING, INC.

MAR 24 1998

Indianapolis Division
6964 Hillsdale Ct.
Indianapolis, IN 46250
Tel: (317) 842-4261
Fax: (317) 842-4286

ANALYTICAL REPORT

Mr. Richard Tyler
MILBANK MANUFACTURING INC
1400 E. HAVENS ST.
KOKOMO, IN 56901-3188

03/19/1998

NET Job Number: 98.01768
Page 1 of 3

Enclosed are the Analytical Results for the following samples submitted to NET, Inc. Indianapolis Division for analysis:

Project Description: TWICW WEEKLY WASTEWATER

Sample Number	Sample Description	Date Taken	Date Received
201729	OUTFALL 001	03/13/1998	03/13/1998

National Environmental Testing, Inc. certifies that the analytical results contained herein apply only to the specific samples analyzed.

Reproduction of this analytical report is permitted only in its entirety.

Len Barrett

Project Representative

MIL0004889



NATIONAL
ENVIRONMENTAL
TESTING, INC.

Indianapolis Division
6964 Hillsdale Ct.
Indianapolis, IN 46250
Tel: (317) 842-4261
Fax: (317) 842-4286

MAR 24 1998

ANALYTICAL REPORT

Mr. Richard Tyler
MILBANK MANUFACTURING INC
1400 E. HAVENS ST.
KOKOMO, IN 56901-3188

03/19/1998

Job No.: 98.01768

Page 2 of 3

Date Received: 03/13/1998

Job Description: TWICW WEEKLY WASTEWATER

Sample Number / Sample I.D.	Result	Flag	Sample Date/ Units	Analyst & Date Analyzed	Method	Reporting Limit
201729	OUTFALL 001		03/13/1998			
Iron, ICP	0.36		mg/L	tjg / 03/18/1998	E-200.7	<0.10
Lead, ICP	<0.080		mg/L	tjg / 03/18/1998	E-200.7	<0.080
Silver, ICP	<0.010		mg/L	tjg / 03/18/1998	E-200.7	<0.010
Zinc, ICP	1.5		mg/L	tjg / 03/18/1998	E-200.7	<0.020

MIL0004890



NATIONAL
ENVIRONMENTAL
TESTING, INC.

Indianapolis Division
6964 Hillsdale Ct.
Indianapolis, IN 46250
Tel: (317) 842-4261
Fax: (317) 842-4286

FIELD REPORT

MAR 24 1998

JOB #: 98.01768
CLIENT: MILBANK MFG.
PROJECT: TWICE WEEKLY WASTEWATER
DATE: 3/12/98
SAMPLER(S): AWS

An ISCO model 6700 auto sampler was used in the composite mode of operation. The sampler was equipped with a plastic composite jug, tygon suction line, battery, ice, and strainer.

All reusable equipment is decontaminated withalconox, tap water, 5% nitric acid, and deionized water. New tygon suction tubing was used for the sampler. A stainless steel strainer was also used for the sampling event.

The sampler was set to take a sample every 30 minutes for 24 hours.

Monitoring start 08:00 on 3/12/98
Monitoring end 08:00 on 3/13/98

The composite sample was cloudy gray.

Notes: Samples were preserved, in the field, in the appropriate containers with any required preservative.

MIL0004891



COMPANY M.I. K
ADDRESS Kokomo
PHONE _____ FAX _____
PROJECT NAME/LOCATION 2x weekly wastewater
PROJECT NUMBER _____
PROJECT MANAGER TODD J. JAMES

REPORT TO: Rich & Tyler

INVOICE TO: _____

P.O. NO. _____

NET QUOTE NO. _____

SAMPLED BY Andrew Sargent
(PRINT NAME)

SIGNATURE

SIGNATURE

ANALYSES

To assist us in selecting the proper method

Is this work being conducted for regulatory compliance monitoring? Yes _____ No _____

Is this work being conducted for regulatory enforcement action? Yes _____ No _____

Which regulations apply: RCRA _____ NPDES Wastewater _____
UST _____ Drinking Water _____
Other _____ None _____

COMMENTS

[illegible]

MAR 24 1998

CONDITION OF SAMPLE: BOTTLES INTACT? YES/NO
FIELD FILTERED? YES/NO

COC SEALS PRESENT AND INTACT? YES/NO
VOLATILES FREE OF HEADSPACE? YES/NO

TEMPERATURE UPON RECEIPT: 5.1°C
Bottles supplied by NET? YES / NO

SAMPLE REMAINDER DISPOSAL: RETURN SAMPLE REMAINDER TO CLIENT VIA _____
 I REQUEST NET TO DISPOSE OF ALL SAMPLE REMAINDERS _____

DATE _____

RELINQUISHED BY:	DATE	TIME
	3/3/98	1615

RECEIVED BY: _____

RELINQUISHED BY:

DATE	TIME
3/3/98	16/9

RECEIVED FOR NET B

METHOD OF SHIPMENT

REMARKS: